

VERIFICATION OF CALIBRATION REPORT 1

of Datamaster cdm
State of Alaska

FEB 18 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>A.W. Andersen</u>	ID# <u>2467</u> Date <u>02/15/2010</u>
Agency <u>Seldovia Police Dept.</u>	Phone # <u>(907) 234-7640</u>
Instrument Location <u>Seldovia Police Dept.</u>	
Datamaster cdm S/N <u>130180</u>	
Alco S/N <u>X301530</u>	Target Value <u>.078</u> High Pressure <u>600 PSI</u>
Alco Test Value Average <u>.074</u>	<u>.074</u>
1 st Alco	2 nd Alco
Signature <u>A.W. Andersen</u>	
(OVER)	

8MB
2/18/2010

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of February, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 01/00



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N

130180 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180 ✓

FEBRUARY 15, 2010 ✓

OPERATOR'S NAME:

ANDERSON AW

OPERATOR'S NUMBER: 2467

SUBJECT'S LAST NAME:

SIXTY DAY VERIFICATION

SUBJECT'S FIRST NAME/MI:

N/A

O.L. #: 1234567890

DEPT/AGENCY: SOU1

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .078 ✓

ALCO S/N: X301530 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.62 in

ALCO TARGET	.078	11:21
BLANK TEST	.000	11:22
INTERNAL STANDARD	VERIFIED	11:22
ALCO TU 29.62 in	.074	11:23
BLANK TEST	.000	11:23
SUBJECT SAMPLE	.000	11:25
BLANK TEST	.000	11:25
ALCO TU 29.62 in	.074	11:26
BLANK TEST	.000	11:26

BMB 2/15/2010

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180 ✓

FEBRUARY 15, 2010 ✓

TIME 11:18

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.62 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""##%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`ab cdefghijklmnop
qrstuvwxyz{|}~